



Product Service

# CERTIFICATE

No. Z1 11 07 77786 001

**Holder of Certificate:** **Kartsana Medical GmbH**

Richardstraße 26  
73054 Eisingen/Fils  
GERMANY

**Production Facility(ies):**

26730

**Certification Mark:**



**Product:**

**Stretchers**  
**Medical stretcher for heavy load**

**Model(s):**

**Kartsana Roll In Trage "Jupiter"**  
**(undercarriage TG-880 and medical stretcher MTR-120)**

**Parameters:**

Length:	1965 mm
Width:	585 mm
Height of the stretcher:	230 mm
Weight of the stretcher:	22 kg (without mattress and retaining belt)
Weight of the undercarriage:	35 kg
Load:	250 kg

**Tested according to:** prEN 1865-3:2010

The product was tested on a voluntary basis and complies with the essential requirements. The certification mark shown above can be affixed on the product. It is not permitted to alter the certification mark in any way. In addition the certification holder must not transfer the certificate to third parties. See also notes overleaf.

**Test report no.:** 71385769

**Date,** 2011-07-29 (Torsten Zimmer)

Page 1 of 1

