



Product Service

CERTIFICATE

No. Z1 16 10 26730 005

Holder of Certificate: **Kartsana S.L.**
 C/Narcías Monturiol 34
 08192 Sant Quirze del Vallés
 SPAIN

Production Facility(ies): 26730

Certification Mark:



Product: **Stretchers for ambulance**

Model(s): **Type: POWER BRAVA**
Model: TG-1000

Parameters:

Length	1970 mm
Width	570 mm
Weight	
Stretcher with Undercarriage (excluding mattress and patient restraints)	70 kg
Loading capacity	300 kg
Nominal voltage	24 VDC
Max. power input	770 W

Tested according to: DIN EN 1865-3:2015

The product was tested on a voluntary basis and complies with the essential requirements. The certification mark shown above can be affixed on the product. It is not permitted to alter the certification mark in any way. In addition the certification holder must not transfer the certificate to third parties. See also notes overleaf.

Test report no.: 713065926

Valid until: 2021-10-31

Date, 2016-11-02 (Torsten Zimmer)

Page 1 of 1

