


Supplement to the test documentation for stretcher MTR 120 on  
carriage TG 880 with locker R 450  
of the Kartsana Medical GmbH

The stretcher MTR 120 on carriage TG 880 with locker R 450 of the Kartsana Medical GmbH was tested in 2010 at the ADAC technical center in Landsberg am Lech. The dynamic tests according to DIN EN 1789: 2007 (D), sections 4.5.9. and 5.3 have been performed.

This is to certify that the former test conditions according to sections 4.5.9. of DIN EN 1789: 2007 (D) are identical with those of the new DIN EN1789:2007+A2:2014. Thus the former test documentation dated 20<sup>th</sup> of December 2010 is also valid for the section 4.5.9. of DIN EN1789:2007+A2:2014.

Landsberg 31.10.2016

A blue ink signature of Volker Sandner, consisting of a large, stylized 'V' and 'S'.

Volker Sandner  
Manager active and passive safety

A black ink signature of Johannes Heilmaier, consisting of a stylized 'J' and 'H'.

Johannes Heilmaier  
Manager crash test facility